

Technical Specification For Reverse Osmosis

SUBJECT	FEATURES	REQUIREMENTS	POINTS
System	All should be equipped by functions	L.P.S	*
		H.P.S	*
	All system should fane	Should be on line without any depending to water tank	*
	-Double softener -Food Grade resin -(Automatic or Volumetric automatic)rinsing -Preferably intelligent -Active carbon filter with automatic rinsing	-Double softener	*
		-Food Grade resin	*
		-(Automatic or Volumetric automatic)rinsing	*
		-Preferably intelligent	*
		-Active carbon filter with automatic rinsing	*
	Hydraulic pump	Should be stainless steel	*
Material	memberance	Should be synthetic except cellulose acetate	*
	Softher housing and active carbon filter should be mad of	1)triple / Quadra hull	* ₁
		2) Stain less steel	* ₂
		3)GRP	* ₃
	Reverse osmosis package	-If case of RO system and pre-treatment system are completed as a one case	* ₄
		-If case of RO system and pre-treatment system are not complete, providing Standard certification for softeners and active carbon filters is mandatory	* ₅
Certificate & standard	Treatment water should be compatible with	AAMI guide lines	*
	Guarantee	1 years from instalation	*
	After sale service	10 years	*
	standard	According to EN	*
		93/42/CEE	*
Athers	Number for 30 units Number for 40 units	600-750 L/H at 15`c	*
		TDS \geq 2000 PPM	*
		1000-1250 L/H at15'c	*
		TDS \geq 2000 PPM	*

- ۱- درصورتیکه جنس بدنه سختی گیرها و فیلتر ذغالی از نوع سه لایه و چند لایه باشد ۵ امتیاز اضافی به جمع امتیازات اضافه می گردد
- ۲- درصورتیکه جنس بدنه سختی گیرها و فیلتر ذغالی از نوع استنلس استیل باشد ۳ امتیاز اضافی به جمع امتیازات اضافه می گردد
- ۳- درصورتیکه جنس بدنه سختی گیرها و فیلتر ذغالی از نوع GRP باشد ۲ امتیاز اضافی به جمع امتیازات اضافه می گردد
- ۴- اگر دستگاه بصورت یک پکیج واحد طراحی شده است ۵ امتیاز تعلق می گیرد
- ۵- اگر دستگاه بصورت یک پکیج نیست لازم است برای سختی گیرها و فیلتر کربن اکتیو گواهی استاندارد ارائه گردد در این صورت ۵ امتیاز تعلق می گیرد.

* وجود مغایرت در هر آیتم باعث حذف کالا خواهد شد.